



REPORT

issued by an Accredited Testing Laboratory

Contact person
Kristina Follin
Safety
+46 10 516 52 21
Kristina.Follin@sp.se

Date 2016-09-08 Reference 5P06665-02rev1 Page 1 (16)



Panthera AB
Gunnebogatan 26
163 53 SPÅNGA

Test of manual wheelchair Panthera U3 Light

2016-09-08: In revision 1, referred test report for Resistance to ignition in 4.4.5 was substituted.

Summary

Manual wheelchair called Panthera U3 Light has been tested in accordance with EN 12183:2014.

The test was performed with a 101.8 kg test dummy on Panthera U3 Light size 42. All test results are valid for Panthera U3 Light, size 33-45.

The Panthera U3 Light fulfilled the requirements in all the performed tests for adult people.



Picture 1 Panthera U3 Light

Picture 2 Panthera U3 Light

SP Technical Research Institute of Sweden Safety - Mechanics Research

Performed by

Signed by: Kristina Follin
Reason: I am the author of this document
Date & Time: 2016-09-08 13:52:38 +02:00

Kristina Follin

Examined by

Signed by: Torbjörn Granberg
Reason: I have reviewed this document
Date & Time: 2016-09-09 08:28:11

Torbjörn Granberg

SP Technical Research Institute of Sweden

Postal address

SP
Box 857
SE-501 15 BORÅS
Sweden

Office location

Brinellgatan 4
SE-504 62 BORÅS

Phone / Fax / E-mail

+46 10 516 50 00
+46 33 13 55 02
info@sp.se

Laboratories are accredited by the Swedish Board for Accreditation and Conformity Assessment (SWEDAC) under the terms of Swedish legislation. This report may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.