



REPORT

issued by an Accredited Testing Laboratory

Contact person
Kristina Follin
Safety
+46 10 516 52 21
Kristina.Follin@sp.se

Date
2016-09-08

Reference
5P06665-01rev1

Page
1 (16)



Panthera AB
Gunnebogatan 26
163 53 SPÅNGA

Test of manual wheelchair Panthera S3 Swing

2016-09-08: In revision 1, referred test report for Resistance to ignition in 4.4.5 was substituted.

Summary

Manual wheelchair called Panthera S3 Swing has been tested in accordance with EN 12183:2014 also including 7176-19:2008 Wheeled mobility devices for use as seats in motor vehicles.

The test was performed with a 101.8 kg test dummy on Panthera S3 Swing size 50. All test results are valid for Panthera S3 Swing, size 33-50.

The Panthera S3 Swing fulfilled the requirements in all the performed tests for adult people.



Picture 1 Panthera S3 Swing

Picture 2 Panthera S3 Swing

SP Technical Research Institute of Sweden Safety - Mechanics Research

Performed by

Kristina Follin

Signed by: Kristina Follin
Reason: I am the author of this document
Date & Time: 2016-09-08 13:53:09 +02:00

Examined by

Torbjörn Granberg

Signed by: Torbjörn Granberg
Reason: I have reviewed this document
Date & Time: 2016-09-09 08:31:21

Kristina Follin

Torbjörn Granberg

SP Technical Research Institute of Sweden

Postal address

SP
Box 857
SE-501 15 BORÅS
Sweden

Office location

Brinellgatan 4
SE-504 62 BORÅS

Phone / Fax / E-mail

+46 10 516 50 00
+46 33 13 55 02
info@sp.se

Laboratories are accredited by the Swedish Board for Accreditation and Conformity Assessment (SWEDAC) under the terms of Swedish legislation. This report may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.