



# REPORT

issued by an Accredited Testing Laboratory

Contact person  
Kristina Follin  
Safety  
+46 10 516 52 21  
Kristina.Follin@sp.se

Date  
2016-09-08

Reference  
4P08158rev2

Page  
1 (17)



Accred. No. 1002  
Testing  
ISO/IEC 17025

Panthera AB  
Gunnabogatan 26  
163 53 SPÅNGA

## Test of manual wheelchair Panthera S3

2016-09-08: In revision 2, referred test report for Resistance to ignition in 4.4.5 was substituted.

### Summary

Manual wheelchair called Panthera S3 and U3, size 33-42 (maximum 100 kg) and size 45-50 (maximum 150 kg) has been tested in accordance with EN 12183:2014 also including 7176-19:2008 Wheeled mobility devices for use as seats in motor vehicles.

The test was performed with a 150.0 kg test dummy on Panthera S3 size 50. All test results are valid for Panthera S3 and U3, size 33-42 (maximum 100 kg) and size 45-50 (maximum 150 kg).

The Panthera S3 fulfilled the requirements in all the performed tests for adult people.



Picture 1 Panthera S3

Picture 2 Panthera S3

### SP Technical Research Institute of Sweden Safety - Mechanics Research

Performed by

*Kristina Follin*

Kristina Follin

Signed by: Kristina Follin  
Reason: I am the author of this document  
Date & Time: 2016-09-08 13:49:23 +02:00

Examined by

*Torbjörn Granberg*

Torbjörn Granberg

Signed by: Torbjörn Granberg  
Reason: I have reviewed this document  
Date & Time: 2016-09-09 08:14:40

---

#### SP Technical Research Institute of Sweden

Postal address  
SP  
Box 857  
SE-501 15 BORÅS  
Sweden

Office location  
Brinellgatan 4  
SE-504 62 BORÅS

Phone / Fax / E-mail  
+46 10 516 50 00  
+46 33 13 55 02  
info@sp.se

Laboratories are accredited by the Swedish Board for Accreditation and Conformity Assessment (SWEDAC) under the terms of Swedish legislation. This report may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.